

# НПЦ "ЭКО-СЕРВИС" САНКТ-ПЕТЕРБУРГ

## TEST PARAMETERS FOR MINDRAY BS-300

TEST	FULL NAME	METHOD	DIRECTION	UNIT	DECIMAL	PRIM WAVE	SEC WAVE	SAMPLE uL	R1 VOL. uL	R2 VOL. uL	LINE LIMIT %	INCUBAT.	REACTION		SUBSTRATE EXHAUST	ANTIGEN SURPLUS	RESPONS		BLANK REAGT.	R1 BLANK		MIX R. BLANK		LINEA		Standard Conce	FACTOR	SPEED	
													START	END			LOWER	UPPER		LOWER	UPPER	LOWER	UPPER	LOWER	UPPER				
ALB	Albumin	Endpoint	Ascend	g/L	1	630	670	3	300	0	20	0	0	25	0	0	0	0	Yes	0	0	0	0	10	70	50			
ALP	Alcaline Phosphatase DGKC	Kinetic	Ascend	U/L	0	405		5	200	50	20	15	5	20	0	0	0	0	No	0	0	0	0	30	700	No	2757		
ALP Mono	Alcaline Phosphatase DGKC	Kinetic	Ascend	U/L	0	405		5	250	0	20	0	5	12	0	0	0	0	No	0	0	0	0	30	700	No	2757		
ALP	Alcaline Phosphatase IFCC	Kinetic	Ascend	U/L	0	405		5	200	50	20	15	5	20	0	0	0	0	No	0	0	0	0	30	700	No	2757		
ALP Mono	Alcaline Phosphatase IFCC	Kinetic	Ascend	U/L	0	405		5	250	0	20	0	5	12	0	0	0	0	No	0	0	0	0	30	700	No	2757		
ALT	ALAT-GPT	Kinetic	Descend	U/L	1	340		20	160	40	20	15	5	20	0	0	0	0	No	0	0	0	0	0	260	No	1750		
ALT Mono	ALAT-GPT	Kinetic	Descend	U/L	1	340		20	200	0	20	0	7	14	0	0	0	0	No	0	0	0	0	0	260	No	1750		
AMYL	Amylase	Kinetic	Ascend	U/L	0	405		5	200	50	20	15	15	25	0	0	0	0	No	0	0	0	0	28	1070	No	4554		
AMYL Mono	Amylase	Kinetic	Ascend	U/L	0	405		5	250	0	20	0	9	23	0	0	0	0	No	0	0	0	0	28	1070	No	4554		
AST	ASAT-GOT	Kinetic	Descend	U/L	0	340		20	160	40	20	15	5	20	0	0	0	0	No	0	0	0	0	0	260	No	1750		
AST Mono	ASAT-GOT	Kinetic	Descend	U/L	0	340		20	200	0	20	0	3	10	0	0	0	0	No	0	0	0	0	0	260	No	1750		
BILT	Bilirubin Total	Endpoint	Ascend	µmol/L	2	510	700	15	200	40	20	15	-1	25	0	0	0	0	Yes	0	0	0	0	0	513	*			
BILD	Bilirubin Direct	Endpoint	Ascend	µmol/L	2	546	700	20	160	40	20	15	-1	25	0	0	0	0	Yes	0	0	0	0	0	171	*			
CA	Calcium	Endpoint	Ascend	mmol/L	2	578	670	3	150	150	20	15	0	25	0	0	0	0	Yes	0	0	0	0	0	5	2,5			
CHOL	Cholesterol	Endpoint	Ascend	mmol/L	2	510	670	3	300	0	20	15	0	25	0	0	0	0	Yes	0	0	0	0	0	19,4	5,17			
CL	Chloride	Endpoint	Ascend	mmol/L	1	510	670	3	300	0	20	15	0	25	0	0	0	0	Yes	0	0	0	0	80	130	100			
CREA	Creatinin	Fixed-Time	Ascend	µmol/L	0	510		30	150	150	20	15	5	10	0	0	0	0	Yes	0	0	0	0	30	1062	177			
CREA Mono	Creatinin	Fixed-Time	Ascend	µmol/L	0	510		30	300	0	20	25	5	10	0	0	0	0	Yes	0	0	0	0	30	1062	177			
GGT	Gamma Glutamyl	Kinetic	Ascend	U/L	0	405		20	160	40	20	15	5	20	0	0	0	0	No	0	0	0	0	8	230	No	1158		
GGT Mono	Gamma Glutamyl	Kinetic	Ascend	U/L	0	405		20	200	0	20	0	5	20	0	0	0	0	No	0	0	0	0	8	230	No	1158		
GLU	Glucose GOD-POD	Endpoint	Ascend	mmol/L	2	510	670	3	300	0	20	15	0	50	0	0	0	0	Yes	0	0	0	0	0	22	5,55			
GLU	Glucose HK	Endpoint	Ascend	mmol/L	2	340	670	3	300	0	20	15	0	25	0	0	0	0	Yes	0	0	0	0	0	40	5,55			
FE	Iron	Endpoint	Ascend	µmol/L	2	578	670	40	200	50	20	15	-1	25	0	0	0	0	Yes	0	0	0	0	3	179	17,9			
LDH	Lactate Dehydrogenase	Kinetic	Descend	U/L	0	340		3	200	50	20	15	5	20	0	0	0	0	No	0	0	0	0	150	1500	No	13400		
LDH Mono	Lactate Dehydrogenase	Kinetic	Descend	U/L	0	340		3	250	0	20	15	5	12	0	0	0	0	No	0	0	0	0	150	1500	No	13400		
Mg	Magnesium	Endpoint	Ascend	mmol/L	2	510	670	3	300	0	20	0	0	25	0	0	0	0	Yes	0	0	0	0	0	2,05	0,82			
PHOS	Inorganic Phosphorus	Endpoint	Ascend	mmol/L	2	340	670	3	300	0	20	15	0	25	0	0	0	0	Yes	0	0	0	0	0	4,8	1,61			
TP	Total Protein	Endpoint	Ascend	g/L	1	546	670	10	200	0	20	15	0	50	0	0	0	0	Yes	0	0	0	0	0	150	50			
TRIG	Triglycerides	Endpoint	Ascend	mmol/L	2	510	670	3	300	0	20	15	0	25	0	0	0	0	Yes	0	0	0	0	0	11,4	2,28			
UA	Uric Acid	Endpoint	Ascend	µmol/L	0	510	670	5	200	50	20	15	0	25	0	0	0	0	Yes	0	0	0	0	0	1428	357,00			
UREA Mono	Urea	Kinetic	Descend	mmol/L	2	340		3	300	0	20	15	5	10	0	0	0	0	No	0	0	0	0	0	33,3	8,33			

\*- точное значение концентрации калибратора указано на этикетке флакона с калибратором